

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071491.D
 Acq On : 23 Mar 2022 02:36
 Operator : JC\MD
 Sample : N1974-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 G-6-5-9-10

Quant Time: Mar 23 07:14:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	718487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1228839	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1208305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	452155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	480469	48.793	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.580%
35) Dibromofluoromethane	8.016	113	359041	46.302	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.600%
50) Toluene-d8	10.439	98	1567391	50.202	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.400%
62) 4-Bromofluorobenzene	12.727	95	573648	54.259	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.520%
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	37959	7.922	ug/l	98
18) Methyl Acetate	4.863	43	17803	1.841	ug/l #	90
20) Methylene Chloride	5.092	84	11144	1.645	ug/l	96
25) 2-Butanone	7.333	43	8609	1.115	ug/l #	87
43) Isopropyl Acetate	8.557	43	79640	3.115	ug/l #	75
95) Naphthalene	15.504	128	23994	6.108	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071491.D
Acq On : 23 Mar 2022 02:36
Operator : JC\MD
Sample : N1974-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
G-6-5-9-10

Quant Time: Mar 23 07:14:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

