

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071492.D
 Acq On : 23 Mar 2022 03:01
 Operator : JC\MD
 Sample : N1970-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 G-3-7-12-13

Quant Time: Mar 23 07:15:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	697577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1199508	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1159753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	423581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	461490	48.270	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.540%
35) Dibromofluoromethane	8.016	113	350454	46.299	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.600%
50) Toluene-d8	10.439	98	1498169	49.158	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.320%
62) 4-Bromofluorobenzene	12.727	95	550807	53.373	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.740%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	64354	13.834	ug/l	95
18) Methyl Acetate	4.857	43	12407	1.454	ug/l	91
20) Methylene Chloride	5.098	84	10780	1.640	ug/l	93
25) 2-Butanone	7.333	43	11830	1.579	ug/l #	80
95) Naphthalene	15.503	128	8045	5.440	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071492.D
Acq On : 23 Mar 2022 03:01
Operator : JC\MD
Sample : N1970-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
G-3-7-12-13

Quant Time: Mar 23 07:15:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

