

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071494.D
 Acq On : 23 Mar 2022 03:49
 Operator : JC\MD
 Sample : N1970-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 G-4-3-6-7

Quant Time: Mar 23 07:15:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	667789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1172868	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1150981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	424967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	455349	49.752	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.500%
35) Dibromofluoromethane	8.016	113	341758	46.176	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.360%
50) Toluene-d8	10.439	98	1481429	49.713	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.420%
62) 4-Bromofluorobenzene	12.727	95	542861	53.797	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.600%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	65512	14.711	ug/l	94
18) Methyl Acetate	4.857	43	20888	2.200	ug/l	94
43) Isopropyl Acetate	8.557	43	69410	2.844	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071494.D
Acq On : 23 Mar 2022 03:49
Operator : JC\MD
Sample : N1970-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
G-4-3-6-7

Quant Time: Mar 23 07:15:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

