

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071497.D
 Acq On : 23 Mar 2022 05:01
 Operator : JC\MD
 Sample : N1972-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 G-5-7-14-15

Quant Time: Mar 23 07:16:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	770553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1321914	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1256197	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	490670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	515024	48.768	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.540%
35) Dibromofluoromethane	8.022	113	387624	46.468	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.940%
50) Toluene-d8	10.439	98	1650582	49.144	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.280%
62) 4-Bromofluorobenzene	12.727	95	622599	54.743	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	109.480%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	89791	17.474	ug/l	94
18) Methyl Acetate	4.857	43	15650	1.593	ug/l	95
20) Methylene Chloride	5.098	84	10632	1.505	ug/l	88
43) Isopropyl Acetate	8.557	43	96479	3.508	ug/l #	83
95) Naphthalene	15.498	128	2922120	121.811	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071497.D
Acq On : 23 Mar 2022 05:01
Operator : JC\MD
Sample : N1972-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
G-5-7-14-15

Quant Time: Mar 23 07:16:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

