

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071457.D
 Acq On : 22 Mar 2022 10:25
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 23 07:05:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.548	0.511	6.8	83	0.00
3 P	Chloromethane	0.822	0.707	14.0	82	0.00
4 C	Vinyl Chloride	0.672	0.690	-2.7#	90	0.00
5 T	Bromomethane	0.335	0.338	-0.9	83	-0.01
6 T	Chloroethane	0.407	0.412	-1.2	84	0.00
7 T	Trichlorofluoromethane	0.884	0.925	-4.6	93	0.00
8 T	Diethyl Ether	0.378	0.413	-9.3	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.512	0.551	-7.6	95	0.00
10 T	Methyl Iodide	0.723	0.712	1.5	83	0.00
11 T	Tert butyl alcohol	0.135	0.137	-1.5	90	0.00
12 CM	1,1-Dichloroethene	0.501	0.515	-2.8#	90	0.00
13 T	Acrolein	0.142	0.117	17.6	76	0.00
14 T	Allyl chloride	1.090	1.159	-6.3	92	0.00
15 T	Acrylonitrile	0.400	0.425	-6.2	91	0.00
16 T	Acetone	0.333	0.397	-19.2	106	-0.01
17 T	Carbon Disulfide	1.427	1.251	12.3	77	0.00
18 T	Methyl Acetate	0.979	1.012	-3.4	97	0.00
19 T	Methyl tert-butyl Ether	1.882	2.157	-14.6	98	-0.01
20 T	Methylene Chloride	0.661	0.646	2.3	91	0.00
21 T	trans-1,2-Dichloroethene	0.548	0.551	-0.5	88	0.00
22 T	Diisopropyl ether	2.148	2.430	-13.1	97	0.00
23 T	Vinyl Acetate	1.818	2.094	-15.2	95	0.00
24 P	1,1-Dichloroethane	1.135	1.241	-9.3	95	0.00
25 T	2-Butanone	0.537	0.609	-13.4	97	0.00
26 T	2,2-Dichloropropane	0.876	1.020	-16.4	100	0.00
27 T	cis-1,2-Dichloroethene	0.651	0.698	-7.2	93	0.00
28 T	Bromochloromethane	0.505	0.466	7.7	79	0.00
29 T	Tetrahydrofuran	0.372	0.389	-4.6	89	0.00
30 C	Chloroform	1.083	1.191	-10.0#	96	0.00
31 T	Cyclohexane	1.094	1.090	0.4	89	0.00
32 T	1,1,1-Trichloroethane	0.922	1.011	-9.7	94	0.00
33 S	1,2-Dichloroethane-d4	0.685	0.652	4.8	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.316	0.303	4.1	86	0.00
36 T	1,1-Dichloropropene	0.506	0.536	-5.9	91	0.00
37 T	Ethyl Acetate	0.696	0.702	-0.9	86	0.00
38 T	Carbon Tetrachloride	0.489	0.528	-8.0	93	0.00
39 T	Methylcyclohexane	0.595	0.650	-9.2	91	0.00
40 TM	Benzene	1.523	1.637	-7.5	92	0.00
41 T	Methacrylonitrile	0.340	0.366	-7.6	92	0.00
42 TM	1,2-Dichloroethane	0.530	0.597	-12.6	95	0.00
43 T	Isopropyl Acetate	1.040	1.174	-12.9	99	0.00
44 TM	Trichloroethene	0.359	0.391	-8.9	92	0.00
45 C	1,2-Dichloropropane	0.409	0.454	-11.0#	94	0.00
46 T	Dibromomethane	0.252	0.278	-10.3	94	0.00
47 T	Bromodichloromethane	0.509	0.573	-12.6	95	0.00
48 T	Methyl methacrylate	0.482	0.559	-16.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071457.D
 Acq On : 22 Mar 2022 10:25
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 23 07:05:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	89	0.00
50 S	Toluene-d8	1.270	1.190	6.3	82	0.00
51 T	4-Methyl-2-Pentanone	0.646	0.736	-13.9	94	0.00
52 CM	Toluene	0.938	1.036	-10.4#	91	0.00
53 T	t-1,3-Dichloropropene	0.561	0.649	-15.7	95	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.701	-14.4	94	0.00
55 T	1,1,2-Trichloroethane	0.376	0.418	-11.2	96	0.00
56 T	Ethyl methacrylate	0.614	0.715	-16.4	96	0.00
57 T	1,3-Dichloropropane	0.675	0.753	-11.6	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.101	0.150	-48.5#	114	0.00
59 T	2-Hexanone	0.441	0.533	-20.9	97	0.00
60 T	Dibromochloromethane	0.373	0.431	-15.5	94	0.00
61 T	1,2-Dibromoethane	0.384	0.417	-8.6	92	0.00
62 S	4-Bromofluorobenzene	0.430	0.437	-1.6	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.349	0.363	-4.0	89	0.00
65 PM	Chlorobenzene	1.074	1.191	-10.9	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.446	-13.2	94	0.00
67 C	Ethyl Benzene	1.899	2.191	-15.4#	94	0.00
68 T	m/p-Xylenes	0.695	0.821	-18.1	94	0.00
69 T	o-Xylene	0.696	0.822	-18.1	94	0.00
70 T	Styrene	1.078	1.353	-25.5#	98	0.00
71 P	Bromoform	0.299	0.356	-19.1	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	4.820	4.880	-1.2	95	0.00
74 T	N-amyl acetate	1.812	2.037	-12.4	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.655	1.627	1.7	96	0.00
76 T	1,2,3-Trichloropropane	1.407	1.243	11.7	91	0.00
77 T	Bromobenzene	1.124	1.133	-0.8	96	0.00
78 T	n-propylbenzene	5.172	5.596	-8.2	98	0.00
79 T	2-Chlorotoluene	3.320	3.405	-2.6	97	0.00
80 T	1,3,5-Trimethylbenzene	3.750	3.945	-5.2	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.463	0.519	-12.1	101	0.00
82 T	4-Chlorotoluene	2.957	3.250	-9.9	100	0.00
83 T	tert-Butylbenzene	3.347	3.607	-7.8	98	0.00
84 T	1,2,4-Trimethylbenzene	3.592	3.884	-8.1	97	0.00
85 T	sec-Butylbenzene	4.448	4.910	-10.4	98	0.00
86 T	p-Isopropyltoluene	3.498	4.006	-14.5	99	0.00
87 T	1,3-Dichlorobenzene	1.826	1.941	-6.3	100	0.00
88 T	1,4-Dichlorobenzene	1.785	1.869	-4.7	101	0.00
89 T	n-Butylbenzene	2.736	3.125	-14.2	102	0.00
90 T	Hexachloroethane	0.696	0.743	-6.8	98	0.00
91 T	1,2-Dichlorobenzene	1.756	1.868	-6.4	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.280	0.300	-7.1	102	0.00
93 T	1,2,4-Trichlorobenzene	0.797	0.883	-10.8	103	0.00
94 T	Hexachlorobutadiene	0.493	0.528	-7.1	98	0.00
95 T	Naphthalene	2.115	2.565	-21.3	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071457.D
Acq On : 22 Mar 2022 10:25
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 23 07:05:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.775	0.881	-13.7	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6