

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032319\
 Data File : VN054600.D
 Acq On : 23 Mar 2019 13:14
 Operator : JC/SP
 Sample : K2020-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-190318

Quant Time: Mar 25 04:03:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	664539	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	983079	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	809537	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	269279	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	427872	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
35) Dibromofluoromethane	7.59	113	332214	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
50) Toluene-d8	10.09	98	1258954	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
62) 4-Bromofluorobenzene	12.40	95	376273	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032319\
Data File : VN054600.D
Acq On : 23 Mar 2019 13:14
Operator : JC/SP
Sample : K2020-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-190318

Quant Time: Mar 25 04:03:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

