

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032319\
 Data File : VN054609.D
 Acq On : 23 Mar 2019 17:22
 Operator : JC/SP
 Sample : K2020-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KY038PS027-190318

Quant Time: Mar 25 04:24:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	601835	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	962448	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	763253	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	264063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	427754	57.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.28%	
35) Dibromofluoromethane	7.59	113	337879	54.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.98%	
50) Toluene-d8	10.09	98	1242704	53.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.62%	
62) 4-Bromofluorobenzene	12.40	95	364019	45.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.78%	
Target Compounds						
4) Vinyl Chloride	2.19	62	4732	0.602	ug/l	96
27) cis-1,2-Dichloroethene	6.83	96	245801	34.766	ug/l	97
44) Trichloroethene	8.83	130	32528	4.547	ug/l	96
64) Tetrachloroethene	10.63	164	23355	3.689	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032319\
Data File : VN054609.D
Acq On : 23 Mar 2019 17:22
Operator : JC/SP
Sample : K2020-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KY038PS027-190318

Quant Time: Mar 25 04:24:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

