

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032319\
 Data File : VN054611.D
 Acq On : 23 Mar 2019 18:16
 Operator : JC/SP
 Sample : K2020-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KY038PS035-190318

Quant Time: Mar 25 04:28:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	617928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	968699	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	781981	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	267931	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	430189	55.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.94%	
35) Dibromofluoromethane	7.59	113	338026	54.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.32%	
50) Toluene-d8	10.09	98	1262917	53.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.66%	
62) 4-Bromofluorobenzene	12.40	95	375780	47.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032319\
Data File : VN054611.D
Acq On : 23 Mar 2019 18:16
Operator : JC/SP
Sample : K2020-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KY038PS035-190318

Quant Time: Mar 25 04:28:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

