

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032319\
 Data File : VN054614.D
 Acq On : 23 Mar 2019 19:39
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 25 04:40:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.478	0.506	-5.9	86	0.00
3 P	Chloromethane	0.650	0.697	-7.2	88	0.00
4 C	Vinyl Chloride	0.653	0.663	-1.5#	82	0.00
5 T	Bromomethane	0.422	0.371	12.1	77	0.00
6 T	Chloroethane	0.392	0.367	6.4	78	0.00
7 T	Trichlorofluoromethane	0.793	0.810	-2.1	83	0.00
8 T	Diethyl Ether	0.305	0.287	5.9	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.466	0.420	9.9	76	0.00
10 T	Methyl Iodide	0.686	0.635	7.4	74	0.00
11 T	Tert butyl alcohol	0.038	0.043	-13.2	89	0.00
12 CM	1,1-Dichloroethene	0.482	0.420	12.9#	72	0.00
13 T	Acrolein	0.052	0.015	71.2#	24#	0.00
14 T	Allyl chloride	0.908	1.005	-10.7	91	0.00
15 T	Acrylonitrile	0.188	0.220	-17.0	92	0.00
16 T	Acetone	0.187	0.158	15.5	66	0.00
17 T	Carbon Disulfide	1.292	1.283	0.7	84	0.00
18 T	Methyl Acetate	0.417	0.517	-24.0	100	0.00
19 T	Methyl tert-butyl Ether	1.407	1.573	-11.8	90	0.00
20 T	Methylene Chloride	0.551	0.581	-5.4	90	0.00
21 T	trans-1,2-Dichloroethene	0.492	0.535	-8.7	88	0.00
22 T	Diisopropyl ether	1.791	2.122	-18.5	95	0.00
23 T	Vinyl Acetate	1.265	1.563	-23.6	94	0.00
24 P	1,1-Dichloroethane	0.964	1.102	-14.3	92	0.00
25 T	2-Butanone	0.240	0.281	-17.1	88	0.00
26 T	2,2-Dichloropropane	0.688	0.687	0.1	79	0.00
27 T	cis-1,2-Dichloroethene	0.587	0.626	-6.6	88	0.00
28 T	Bromochloromethane	0.475	0.553	-16.4	93	0.00
29 T	Tetrahydrofuran	0.155	0.194	-25.2#	94	0.00
30 C	Chloroform	0.941	1.054	-12.0#	91	0.00
31 T	Cyclohexane	1.121	1.009	10.0	89	0.00
32 T	1,1,1-Trichloroethane	0.777	0.887	-14.2	90	0.00
33 S	1,2-Dichloroethane-d4	0.622	0.697	-12.1	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.322	0.329	-2.2	86	0.00
36 T	1,1-Dichloropropene	0.468	0.466	0.4	88	0.00
37 T	Ethyl Acetate	0.333	0.376	-12.9	94	0.00
38 T	Carbon Tetrachloride	0.454	0.453	0.2	88	0.00
39 T	Methylcyclohexane	0.535	0.492	8.0	81	0.00
40 TM	Benzene	1.337	1.386	-3.7	90	0.00
41 T	Methacrylonitrile	0.203	0.200	1.5	87	0.00
42 TM	1,2-Dichloroethane	0.487	0.512	-5.1	93	0.00
43 T	Isopropyl Acetate	0.562	0.664	-18.1	97	0.00
44 TM	Trichloroethene	0.372	0.359	3.5	85	0.00
45 C	1,2-Dichloropropane	0.359	0.397	-10.6#	93	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032319\
 Data File : VN054614.D
 Acq On : 23 Mar 2019 19:39
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 25 04:40:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.228	-6.5	91	0.00
47 T	Bromodichloromethane	0.444	0.482	-8.6	92	0.00
48 T	Methyl methacrylate	0.305	0.336	-10.2	94	0.00
49 T	1,4-Dioxane	0.003	0.004	-33.3#	86	0.00
50 S	Toluene-d8	1.211	1.197	1.2	84	0.00
51 T	4-Methyl-2-Pentanone	0.311	0.367	-18.0	96	0.00
52 CM	Toluene	0.822	0.825	-0.4#	87	0.00
53 T	t-1,3-Dichloropropene	0.414	0.466	-12.6	90	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.555	-8.8	90	0.00
55 T	1,1,2-Trichloroethane	0.296	0.313	-5.7	90	0.00
56 T	Ethyl methacrylate	0.397	0.446	-12.3	92	0.00
57 T	1,3-Dichloropropane	0.514	0.552	-7.4	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.163	0.167	-2.5	83	0.00
59 T	2-Hexanone	0.203	0.232	-14.3	93	0.00
60 T	Dibromochloromethane	0.315	0.347	-10.2	90	0.00
61 T	1,2-Dibromoethane	0.292	0.307	-5.1	88	0.00
62 S	4-Bromofluorobenzene	0.412	0.414	-0.5	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.415	0.371	10.6	79	0.00
65 PM	Chlorobenzene	1.017	1.021	-0.4	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.394	-7.9	91	0.00
67 C	Ethyl Benzene	1.796	1.821	-1.4#	88	0.00
68 T	m/p-Xylenes	0.680	0.669	1.6	86	0.00
69 T	o-Xylene	0.664	0.664	0.0	87	0.00
70 T	Styrene	1.079	1.105	-2.4	87	0.00
71 P	Bromoform	0.222	0.247	-11.3	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	4.029	3.882	3.6	87	0.00
74 T	N-amyl acetate	1.159	1.351	-16.6	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.920	1.019	-10.8	94	0.00
76 T	1,2,3-Trichloropropane	0.793	0.851	-7.3	92	0.00
77 T	Bromobenzene	0.999	0.981	1.8	87	0.00
78 T	n-propylbenzene	4.369	4.267	2.3	86	0.00
79 T	2-Chlorotoluene	2.724	2.634	3.3	87	0.00
80 T	1,3,5-Trimethylbenzene	3.302	3.186	3.5	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.205	0.228	-11.2	90	0.00
82 T	4-Chlorotoluene	2.641	2.613	1.1	87	0.00
83 T	tert-Butylbenzene	2.874	2.734	4.9	84	0.00
84 T	1,2,4-Trimethylbenzene	3.240	3.105	4.2	83	0.00
85 T	sec-Butylbenzene	3.583	3.415	4.7	85	0.00
86 T	p-Isopropyltoluene	3.206	3.050	4.9	83	0.00
87 T	1,3-Dichlorobenzene	1.656	1.602	3.3	83	0.00
88 T	1,4-Dichlorobenzene	1.611	1.565	2.9	83	0.00
89 T	n-Butylbenzene	2.507	2.423	3.4	80	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032319\
Data File : VN054614.D
Acq On : 23 Mar 2019 19:39
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 25 04:40:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.498	0.526	-5.6	89	0.00
91 T	1,2-Dichlorobenzene	1.604	1.596	0.5	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.135	0.158	-17.0	89	0.00
93 T	1,2,4-Trichlorobenzene	0.800	0.781	2.4	77	0.00
94 T	Hexachlorobutadiene	0.523	0.434	17.0	76	0.00
95 T	Naphthalene	1.783	1.832	-2.7	80	0.00
96 T	1,2,3-Trichlorobenzene	0.796	0.789	0.9	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6