


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN032319\
 Data File : VN054595.D
 Acq On : 23 Mar 2019 9:47
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Title : SW846 8260
 Last Update : Fri Mar 22 02:09:07 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	16412	PASS
75	95	30	60	53.0	42128	PASS
95	95	100	100	100.0	79450	PASS
96	95	5	9	7.2	5730	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	83.4	66296	PASS
175	174	5	9	7.8	5143	PASS
176	174	95	101	98.6	65394	PASS
177	176	5	9	6.5	4256	PASS