

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032419\
 Data File : VN054639.D
 Acq On : 24 Mar 2019 11:09
 Operator : JC/SP
 Sample : VN0324MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0324MBL01

Quant Time: Mar 25 06:45:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	620422	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	947851	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	792527	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	287229	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	390853	50.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.30%	
35) Dibromofluoromethane	7.59	113	299083	48.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.96%	
50) Toluene-d8	10.09	98	1135782	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	
62) 4-Bromofluorobenzene	12.40	95	352725	45.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.30%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032419\
Data File : VN054639.D
Acq On : 24 Mar 2019 11:09
Operator : JC/SP
Sample : VN0324MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0324MBL01

Quant Time: Mar 25 06:45:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

