

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032419\
 Data File : VN054641.D
 Acq On : 24 Mar 2019 12:10
 Operator : JC/SP
 Sample : K2047-01ME
 Misc : 6.05g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022SB432-190320-(26-28)ME

Quant Time: Mar 25 06:51:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	517547	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	828213	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	719224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	290544	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	330171	51.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.58%	
35) Dibromofluoromethane	7.59	113	252104	47.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.50%	
50) Toluene-d8	10.09	98	998560	49.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.56%	
62) 4-Bromofluorobenzene	12.40	95	325844	47.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.48%	
Target Compounds						
4) Vinyl Chloride	2.19	62	49061	7.259	ug/l	100
27) cis-1,2-Dichloroethene	6.82	96	78704	12.945	ug/l	98
44) Trichloroethene	8.83	130	20872	3.390	ug/l	82
52) Toluene	10.15	92	41566	3.052	ug/l	99
65) Chlorobenzene	11.43	112	92364	6.314	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032419\
Data File : VN054641.D
Acq On : 24 Mar 2019 12:10
Operator : JC/SP
Sample : K2047-01ME
Misc : 6.05g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022SB432-190320-(26-28)ME

Quant Time: Mar 25 06:51:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

