

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032419\
 Data File : VN054646.D
 Acq On : 24 Mar 2019 14:22
 Operator : JC/SP
 Sample : K2047-08
 Misc : 6.55g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022SB445-190320-FD-(17-19)

Quant Time: Mar 25 09:36:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	552391	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	871905	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	742667	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	315131	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	346617	50.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.90%	
35) Dibromofluoromethane	7.59	113	261142	46.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.98%	
50) Toluene-d8	10.09	98	1061487	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
62) 4-Bromofluorobenzene	12.40	95	337793	47.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.02%	
Target Compounds						
39) Methylcyclohexane	9.08	83	41848	4.488	ug/l #	81
65) Chlorobenzene	11.43	112	709143	46.948	ug/l	99
85) sec-Butylbenzene	13.17	105	30614	1.356	ug/l	100
87) 1,3-Dichlorobenzene	13.28	146	20295	1.945	ug/l #	67
88) 1,4-Dichlorobenzene	13.36	146	44765	4.408	ug/l #	80
89) n-Butylbenzene	13.61	91	27836	1.762	ug/l	91
91) 1,2-Dichlorobenzene	13.65	146	60328	5.968	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032419\
Data File : VN054646.D
Acq On : 24 Mar 2019 14:22
Operator : JC/SP
Sample : K2047-08
Misc : 6.55uL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022SB445-190320-FD-(17-19)

Quant Time: Mar 25 09:36:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

