

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032420\
 Data File : VN060674.D
 Acq On : 24 Mar 2020 15:59
 Operator : JC/MD
 Sample : PB127723ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127723ZHE#02

Quant Time: Mar 25 09:57:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	190131	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	315397	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	292940	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	132830	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	124287	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
35) Dibromofluoromethane	7.56	113	92870	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
50) Toluene-d8	10.08	98	388215	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
62) 4-Bromofluorobenzene	12.40	95	142033	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
Target Compounds						
16) Acetone	3.78	43	4869	5.266	ug/l	99
20) Methylene Chloride	4.52	84	5095	2.240	ug/l	95
43) Isopropyl Acetate	8.14	43	53734	9.874	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032420\
Data File : VN060674.D
Acq On : 24 Mar 2020 15:59
Operator : JC/MD
Sample : PB127723ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127723ZHE#02

Quant Time: Mar 25 09:57:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

