

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032420\
 Data File : VN060675.D
 Acq On : 24 Mar 2020 16:22
 Operator : JC/MD
 Sample : PB127723ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127723ZHE#03

Quant Time: Mar 25 10:00:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	198636	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	330854	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	308384	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	140527	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	127699	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
35) Dibromofluoromethane	7.57	113	96279	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
50) Toluene-d8	10.08	98	407052	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
62) 4-Bromofluorobenzene	12.40	95	150894	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
Target Compounds						
16) Acetone	3.79	43	4931	5.105	ug/l	95
20) Methylene Chloride	4.53	84	4780	2.012	ug/l	97
43) Isopropyl Acetate	8.14	43	53776	9.420	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032420\
Data File : VN060675.D
Acq On : 24 Mar 2020 16:22
Operator : JC/MD
Sample : PB127723ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127723ZHE#03

Quant Time: Mar 25 10:00:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

