

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032420\
 Data File : VN060681.D
 Acq On : 24 Mar 2020 18:38
 Operator : JC/MD
 Sample : L2006-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OIL-WATER-SEPERATOR

Manual Integrations
 APPROVED

MMDadoda
 3/25/2020 10:20:41 PM

Quant Time: Mar 25 10:11:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.85	168	167476	50.00	ug/l	0.21
34) 1,4-Difluorobenzene	8.71	114	286097	50.00	ug/l	0.14
63) Chlorobenzene-d5	11.43	117	255406	50.00	ug/l	0.03
72) 1,4-Dichlorobenzene-d4	13.35	152	109616	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.20	65	107408	46.72	ug/l	0.19
Spiked Amount	50.000		Recovery	=	93.44%	
35) Dibromofluoromethane	7.79	113	86565	50.04	ug/l	0.23
Spiked Amount	50.000		Recovery	=	100.08%	
50) Toluene-d8	10.15	98	343322	48.59	ug/l	0.07
Spiked Amount	50.000		Recovery	=	97.18%	
62) 4-Bromofluorobenzene	12.41	95	118587	45.79	ug/l	0.02
Spiked Amount	50.000		Recovery	=	91.58%	
Target Compounds						
16) Acetone	4.33	43	39744m	48.803	ug/l	Qvalue
20) Methylene Chloride	5.06	84	9209m	4.597	ug/l	
25) 2-Butanone	7.14	43	12567m	9.536	ug/l	
43) Isopropyl Acetate	8.31	43	36206	7.334	ug/l #	89
86) p-Isopropyltoluene	13.29	119	21868	3.125	ug/l	98
95) Naphthalene	15.13	128	7821	1.267	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032420\
Data File : VN060681.D
Acq On : 24 Mar 2020 18:38
Operator : JC/MD
Sample : L2006-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OIL-WATER-SEPERATOR

Manual Integrations
APPROVED

MMDadoda
3/25/2020 10:20:41 PM

Quant Time: Mar 25 10:11:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

