

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032420\
 Data File : VN060682.D
 Acq On : 24 Mar 2020 19:01
 Operator : JC/MD
 Sample : L2011-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GHD1-WC-S-BC-IBC

Quant Time: Mar 25 09:54:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	194465	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	325896	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	303443	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	143631	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	127574	47.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.58%	
35) Dibromofluoromethane	7.56	113	99177	50.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.66%	
50) Toluene-d8	10.08	98	400577	49.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.54%	
62) 4-Bromofluorobenzene	12.40	95	149923	50.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.64%	
Target Compounds						
16) Acetone	3.79	43	16169	17.099	ug/l	93
20) Methylene Chloride	4.53	84	6403	2.753	ug/l	97
25) 2-Butanone	6.80	43	3343	2.185	ug/l	91
43) Isopropyl Acetate	8.14	43	44581	7.928	ug/l #	91
54) cis-1,3-Dichloropropene	9.89	75	19107	4.801	ug/l #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032420\
Data File : VN060682.D
Acq On : 24 Mar 2020 19:01
Operator : JC/MD
Sample : L2011-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GHD1-WC-S-BC-IBC

Quant Time: Mar 25 09:54:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

