

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032420\
 Data File : VN060683.D
 Acq On : 24 Mar 2020 19:24
 Operator : JC/MD
 Sample : L2011-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GHD1-WC-S-SB-01

Quant Time: Mar 25 09:38:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	190237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	317671	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	302927	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145243	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	124936	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
35) Dibromofluoromethane	7.56	113	95581	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
50) Toluene-d8	10.08	98	395604	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
62) 4-Bromofluorobenzene	12.40	95	151357	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
Target Compounds						
16) Acetone	3.78	43	10938	11.824	ug/l	99
18) Methyl Acetate	4.29	43	2782	1.068	ug/l #	66
20) Methylene Chloride	4.52	84	4620	2.030	ug/l	99
43) Isopropyl Acetate	8.14	43	60391	11.017	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032420\
Data File : VN060683.D
Acq On : 24 Mar 2020 19:24
Operator : JC/MD
Sample : L2011-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GHD1-WC-S-SB-01

Quant Time: Mar 25 09:38:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

