

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032421\
 Data File : VN066315.D
 Acq On : 24 Mar 2021 18:16
 Operator : JC/MD
 Sample : VN0324WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0324WBL01

Quant Time: Mar 25 06:00:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 25 05:52:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.102	128	81831	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.513	114	444662	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.347	117	373055	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.944	65	187333	29.78	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.27%
60) 4-Bromofluorobenzene	12.345	95	153293	26.33	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	87.77%
63) Toluene-d8	10.025	98	538661	30.77	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.57%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032421\
Data File : VN066315.D
Acq On : 24 Mar 2021 18:16
Operator : JC/MD
Sample : VN0324WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0324WBL01

Quant Time: Mar 25 06:00:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 25 05:52:50 2021
Response via : Initial Calibration

