

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032421\
 Data File : VN066311.D
 Acq On : 24 Mar 2021 14:52
 Operator : JC/MD
 Sample : VSTDICCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020

Manual Integrations
 APPROVED

MMDadoda
 3/30/2021 10:41:08 AM

Quant Time: Mar 25 04:38:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 25 04:35:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.107	128	79987	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.515	114	442629	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.350	117	400913	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.952	65	187035	27.47	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	91.57%		
60) 4-Bromofluorobenzene	12.345	95	187495	27.40	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	91.33%		
63) Toluene-d8	10.028	98	557940	30.82	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	102.73%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.818	85	88914	21.68	ug/l	100
3) Chloromethane	2.017	50	110750	20.52	ug/l	100
4) Vinyl Chloride	2.145	62	117526	20.70	ug/l	100
5) Bromomethane	2.510	94	80243	20.41	ug/l	100
6) Chloroethane	2.647	64	74580	22.00	ug/l	100
7) Trichlorofluoromethane	2.961	101	164931	19.41	ug/l	100
8) Diethyl Ether	3.342	74	57695	19.87	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.685	101	82863	22.47	ug/l	100
10) 1,1-Dichloroethene	3.661	96	79907	19.74	ug/l	100
11) Methyl Iodide	3.875	142	121805	19.36	ug/l	100
12) Methyl Acetate	4.235	43	133118	27.69	ug/l	100
13) Acrolein	3.535	56	36974	54.14	ug/l	100
14) Acrylonitrile	4.876	53	266653	119.74	ug/l	100
15) Acetone	3.736	58	72135	113.84	ug/l	100
16) Carbon Disulfide	3.974	76	226904	19.18	ug/l	100
17) Allyl chloride	4.235	41	158753	21.27	ug/l	100
18) Methylene Chloride	4.457	84	98848	19.74	ug/l	100
19) trans-1,2-Dichloroethene	4.943	96	89448	19.47	ug/l	100
20) Diisopropyl ether	5.855	45	328413	22.38	ug/l	100
21) 1,1-Dichloroethane	5.750	63	178189	19.90	ug/l	100
22) cis-1,2-Dichloroethene	6.734	96	108449	19.35	ug/l	100
23) tert-Butyl Alcohol	4.680	59	113563	116.12	ug/l #	100
24) Methyl tert-Butyl Ether	4.937	73	314886	19.29	ug/l	100
25) Chloroform	7.287	83	180442	18.84	ug/l	100
26) Cyclohexane	7.574	56	156685	22.18	ug/l #	100
29) 1,1-Dichloropropene	7.713	75	128493	19.48	ug/l	100
30) 2-Butanone	6.734	43	379656	131.16	ug/l	100
31) 2,2-Dichloropropane	6.729	77	160886	18.85	ug/l	100
32) 1,1,1-Trichloroethane	7.491	97	155973	18.52	ug/l	100
33) Carbon Tetrachloride	7.697	117	132663	18.39	ug/l	100
34) Benzene	7.965	78	397407	19.82	ug/l	100
35) Methacrylonitrile	7.080	41	61092	19.99	ug/l	100
36) 1,2-Dichloroethane	8.048	62	148014	18.40	ug/l	100
37) Trichloroethene	8.767	130	102569	18.99	ug/l	100
38) Methylcyclohexane	9.014	83	158876	21.76	ug/l	100
39) 1,2-Dichloropropane	9.051	63	107590	20.80	ug/l	100
40) Dibromomethane	9.143	93	69071	19.83	ug/l	100
41) Bromodichloromethane	9.338	83	146494	18.87	ug/l	100
42) Vinyl Acetate	5.793	43	1438940	115.59	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032421\
 Data File : VN066311.D
 Acq On : 24 Mar 2021 14:52
 Operator : JC/MD
 Sample : VSTDICCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020

Manual Integrations
 APPROVED

MMDadoda
 3/30/2021 10:41:08 AM

Quant Time: Mar 25 04:38:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 25 04:35:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	6.836	43	150591	23.87	ug/l #	100
44) Isopropyl Acetate	8.089	43	266113	24.46	ug/l #	100
45) 1,4-Dioxane	9.132	88	34087	387.91	ug/l	100
46) Methyl methacrylate	9.135	41	122820	24.19	ug/l	100
47) n-amyl Acetate	12.020	43	112430	15.63	ug/l	100
48) t-1,3-Dichloropropene	10.323	75	158248	17.83	ug/l	100
49) cis-1,3-Dichloropropene	9.776	75	170114	18.47	ug/l	100
50) 1,1,2-Trichloroethane	10.502	97	100026	20.27	ug/l	100
51) Ethyl methacrylate	10.371	69	156358	20.69	ug/l	100
52) 1,3-Dichloropropane	10.647	76	166100	19.46	ug/l	100
53) Dibromochloromethane	10.843	129	112532	18.90	ug/l	100
54) 1,2-Dibromoethane	10.948	107	103188	19.91	ug/l	100
55) 2-Chloroethyl vinyl ether	9.631	63	121867	52.77	ug/l	100
56) Bromoform	12.071	173	82050	20.41	ug/l #	100
58) 4-Methyl-2-Pentanone	9.920	43	791312	134.55	ug/l	100
59) 2-Hexanone	10.693	43	544697	132.41	ug/l	100
61) Tetrachloroethene	10.572	164	108321	20.65	ug/l	100
62) Toluene	10.092	91	429887	20.27	ug/l	100
64) Chlorobenzene	11.377	112	255842	18.84	ug/l	100
65) 1,1,1,2-Tetrachloroethane	11.452	131	99655	19.06	ug/l	100
66) Ethyl Benzene	11.455	91	466172	19.35	ug/l	100
67) m/p-Xylenes	11.564	106	345684	37.29	ug/l	100
68) o-Xylene	11.892	106	171062	18.77	ug/l	100
69) Styrene	11.908	104	275409	18.41	ug/l	100
70) Isopropylbenzene	12.192	105	452539	18.95	ug/l	100
71) 1,1,2,2-Tetrachloroethane	12.447	83	148850	22.17	ug/l	100
72) 1,2,3-Trichloropropane	12.498	75	144643m	21.60	ug/l	
73) Bromobenzene	12.468	156	112514	19.41	ug/l	100
74) n-propylbenzene	12.535	91	497879	18.83	ug/l	100
75) 2-Chlorotoluene	12.618	91	301524	18.03	ug/l	100
76) 1,3,5-Trimethylbenzene	12.677	105	370009	17.53	ug/l	100
77) t-1,4-Dichloro-2-butene	12.246	75	46984	19.95	ug/l	100
78) 4-Chlorotoluene	12.718	91	299417	17.64	ug/l	100
79) tert-butylbenzene	12.938	119	316842	17.99	ug/l	100
80) 1,2,4-Trimethylbenzene	12.983	105	354677	17.13	ug/l	100
81) sec-Butylbenzene	13.115	105	410254	18.46	ug/l	100
82) p-Isopropyltoluene	13.233	119	358212	17.21	ug/l	100
83) 1,3-Dichlorobenzene	13.227	146	181171	18.24	ug/l	100
84) 1,4-Dichlorobenzene	13.308	146	178841	18.52	ug/l	100
85) n-Butylbenzene	13.560	91	293028	17.60	ug/l	100
86) Hexachloroethane	13.815	117	66901	19.71	ug/l	100
87) 1,2-Dichlorobenzene	13.597	146	180093	18.38	ug/l	100
88) 1,2-Dibromo-3-Chloropr...	14.214	75	30412	19.36	ug/l	100
89) 1,2,4-Trichlorobenzene	14.850	180	111598	20.08	ug/l	100
90) Hexachlorobutadiene	14.949	225	59897	23.46	ug/l	100
91) Naphthalene	15.070	128	346332	21.16	ug/l	100
92) 1,2,3-Trichlorobenzene	15.249	180	110621	19.81	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032421\
Data File : VN066311.D
Acq On : 24 Mar 2021 14:52
Operator : JC/MD
Sample : VSTDICCC020
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED

MMDadoda
3/30/2021 10:41:08 AM

Quant Time: Mar 25 04:38:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 25 04:35:34 2021
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.00 ug/l
RT: 7.107 min Scan# 2020
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:128 Resp: 79987
Ion Ratio Lower Upper
128 100
49 171.7 0.0 429.3
130 131.4 0.0 328.5

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#2
Dichlorodifluoromethane
Concen: 21.68 ug/l
RT: 1.818 min Scan# 48
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 85 Resp: 88914
Ion Ratio Lower Upper
85 100
87 30.9 15.4 46.4









#3
Chloromethane
Concen: 20.52 ug/l
RT: 2.017 min Scan# 122
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 50 Resp: 110750
Ion Ratio Lower Upper
50 100
52 32.8 26.2 39.4

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#4
Vinyl Chloride
Concen: 20.70 ug/l
RT: 2.145 min Scan# 170
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Tgt Ion: 62 Resp: 117526
Ion Ratio Lower Upper
62 100
64 32.2 25.8 38.6








#5
Bromomethane
Concen: 20.41 ug/l
RT: 2.510 min Scan# 306
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020





Tgt Ion: 94 Resp: 80243
Ion Ratio Lower Upper
94 100
96 97.3 77.8 116.8




#6
Chloroethane
Concen: 22.00 ug/l
RT: 2.647 min Scan# 357
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52



Tgt Ion: 64 Resp: 74580
Ion Ratio Lower Upper
64 100
66 30.3 24.2 36.4









#7
Trichlorofluoromethane
Concen: 19.41 ug/l
RT: 2.961 min Scan# 474
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 101 Resp: 164931
Ion Ratio Lower Upper
101 100
103 64.5 51.6 77.4

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#8
Diethyl Ether
Concen: 19.87 ug/l
RT: 3.342 min Scan# 616
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 74 Resp: 57695
Ion Ratio Lower Upper
74 100
45 98.0 19.5 175.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 22.47 ug/l
RT: 3.685 min Scan# 744
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 101 Resp: 82863
Ion Ratio Lower Upper
101 100
85 45.3 0.0 90.6
151 79.5 0.0 159.0

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#10
1,1-Dichloroethene
Concen: 19.74 ug/l
RT: 3.661 min Scan# 735
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Tgt Ion: 96 Resp: 79907
Ion Ratio Lower Upper
96 100
61 174.5 139.6 209.4
98 67.1 53.7 80.5








#11
Methyl Iodide
Concen: 19.36 ug/l
RT: 3.875 min Scan# 815
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020





Tgt Ion:142 Resp: 121805
Ion Ratio Lower Upper
142 100
127 41.1 20.5 61.6
141 14.9 7.4 22.3




#12
Methyl Acetate
Concen: 27.69 ug/l
RT: 4.235 min Scan# 949
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52



Tgt Ion: 43 Resp: 133118
Ion Ratio Lower Upper
43 100
74 22.5 18.0 27.0









#13
Acrolein
Concen: 54.14 ug/l
RT: 3.535 min Scan# 688
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 56 Resp: 36974
Ion Ratio Lower Upper
56 100
55 67.6 54.1 81.1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM




#14
Acrylonitrile
Concen: 119.74 ug/l
RT: 4.876 min Scan# 1188
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 53 Resp: 266653
Ion Ratio Lower Upper
53 100
52 82.0 41.0 123.0
51 35.6 17.8 53.4






#15
Acetone
Concen: 113.84 ug/l
RT: 3.736 min Scan# 763
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52





Tgt Ion: 58 Resp: 72135
Ion Ratio Lower Upper
58 100
43 334.1 267.3 400.9




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#16
Carbon Disulfide
Concen: 19.18 ug/l
RT: 3.974 min Scan# 852
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Tgt Ion: 76 Resp: 226904
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.6





#17
Allyl chloride
Concen: 21.27 ug/l
RT: 4.235 min Scan# 949
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampled :
VSTDICCC020



Tgt Ion: 41 Resp: 158753
Ion Ratio Lower Upper
41 100
39 77.3 38.6 116.0
76 35.2 17.6 52.8




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#18
Methylene Chloride
Concen: 19.74 ug/l
RT: 4.457 min Scan# 1032
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 84 Resp: 98848
Ion Ratio Lower Upper
84 100
49 123.6 98.9 148.3
51 38.5 30.8 46.2
86 63.7 51.0 76.4





#19





trans-1,2-Dichloroethene
Concen: 19.47 ug/l
RT: 4.943 min Scan# 1213
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 96 Resp: 89448
Ion Ratio Lower Upper
96 100
61 141.6 113.4 170.0
98 64.3 51.0 76.4

Instrument :
MSVOA_N
ClientSampled :
VSTDICCC020

Manual Integrations
APPROVED




MMDadoda
3/30/2021 10:41:08 AM




#20


Diisopropyl ether
Concen: 22.38 ug/l
RT: 5.855 min Scan# 1553
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 45 Resp: 328413
Ion Ratio Lower Upper
45 100
43 64.3 51.4 77.2
87 26.7 21.4 32.0
59 11.5 9.2 13.8






#21
1,1-Dichloroethane
Concen: 19.90 ug/l
RT: 5.750 min Scan# 1514
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52





Tgt Ion: 63 Resp: 178189
Ion Ratio Lower Upper
63 100
98 6.6 3.3 9.9
100 3.8 1.9 5.7




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#22
cis-1,2-Dichloroethene
Concen: 19.35 ug/l
RT: 6.734 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52



Tgt Ion: 96 Resp: 108449
Ion Ratio Lower Upper
96 100
61 149.2 119.4 179.0
98 64.4 51.5 77.3






#23
tert-Butyl Alcohol
Concen: 116.12 ug/l
RT: 4.680 min Scan# 1115
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED

MMDadoda
3/30/2021 10:41:08 AM







Tgt Ion: 59 Resp: 113563
Ion Ratio Lower Upper
59 100
52 0.1 0.1 0.1#



#24
Methyl tert-Butyl Ether
Concen: 19.29 ug/l
RT: 4.937 min Scan# 1211
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 73 Resp: 314886
Ion Ratio Lower Upper
73 100
43 23.1 18.5 27.7









#25
Chloroform
Concen: 18.84 ug/l
RT: 7.287 min Scan# 2087
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 83 Resp: 180442
Ion Ratio Lower Upper
83 100
85 61.8 49.4 74.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#26
Cyclohexane
Concen: 22.18 ug/l
RT: 7.574 min Scan# 2194
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Tgt Ion: 56 Resp: 156685
Ion Ratio Lower Upper
56 100
89 0.0 0.0 0.0
84 81.8 65.4 98.2





#27
1,2-Dichloroethane-d4
Concen: 27.47 ug/l
RT: 7.952 min Scan# 2335
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020



Tgt Ion: 65 Resp: 187035
Ion Ratio Lower Upper
65 100
67 52.8 42.2 63.4




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#28
1,4-Difluorobenzene
Concen: 30.00 ug/l
RT: 8.515 min Scan# 2545
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 114 Resp: 442629
Ion Ratio Lower Upper
114 100
63 21.7 17.4 26.0
88 15.8 12.6 19.0









#29
1,1-Dichloropropene
Concen: 19.48 ug/l
RT: 7.713 min Scan# 2246
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:	75	Resp:	128493
Ion	Ratio	Lower	Upper
75	100		
110	34.7	0.0	69.4
77	30.3	0.0	60.6

Instrument :
MSVOA_N
ClientSampled :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM




#30
2-Butanone
Concen: 131.16 ug/l
RT: 6.734 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:	43	Resp:	379656
Ion	Ratio	Lower	Upper
43	100		
57	8.6	6.9	10.3
72	26.1	20.9	31.3






#31
2,2-Dichloropropane
Concen: 18.85 ug/l
RT: 6.729 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52





Tgt Ion: 77 Resp: 160886
Ion Ratio Lower Upper
77 100
97 22.6 18.1 27.1




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#32
1,1,1-Trichloroethane
Concen: 18.52 ug/l
RT: 7.491 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Tgt Ion: 97 Resp: 155973
Ion Ratio Lower Upper
97 100
99 63.4 50.7 76.1
61 47.0 37.6 56.4





#33
Carbon Tetrachloride
Concen: 18.39 ug/l
RT: 7.697 min Scan# 2240
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Tgt Ion:117 Resp: 132663
Ion Ratio Lower Upper
117 100
119 90.7 72.6 108.8
121 29.4 23.5 35.3

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#34
Benzene
Concen: 19.82 ug/l
RT: 7.965 min Scan# 2340
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52



Tgt Ion: 78 Resp: 397407
Ion Ratio Lower Upper
78 100
77 23.2 18.6 27.8
51 27.4 21.9 32.9









#35
Methacrylonitrile
Concen: 19.99 ug/l
RT: 7.080 min Scan# 2010
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:	41	Resp:	61092
Ion Ratio	Lower	Upper	
41	100		
39	54.5	27.3	81.8
67	62.9	31.4	94.3
52	26.5	13.3	39.8

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#36
1,2-Dichloroethane
Concen: 18.40 ug/l
RT: 8.048 min Scan# 2371
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:	62	Resp:	148014
Ion Ratio	Lower	Upper	
62	100		
98	8.6	6.9	10.3





#37





Trichloroethene
Concen: 18.99 ug/l
RT: 8.767 min Scan# 2639
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:130 Resp: 102569
Ion Ratio Lower Upper
130 100
95 100.9 80.7 121.1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED




MMDadoda
3/30/2021 10:41:08 AM




#38

Methylcyclohexane
Concen: 21.76 ug/l
RT: 9.014 min Scan# 2731
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 83 Resp: 158876
Ion Ratio Lower Upper
83 100
55 77.5 38.8 116.3
98 44.4 22.2 66.6





#39





1,2-Dichloropropane
Concen: 20.80 ug/l
RT: 9.051 min Scan# 2745
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 63 Resp: 107590
Ion Ratio Lower Upper
63 100
65 31.1 24.9 37.3

Instrument :
MSVOA_N
ClientSampled :
VSTDICCC020

Manual Integrations
APPROVED




MMDadoda
3/30/2021 10:41:08 AM




#40

Dibromomethane
Concen: 19.83 ug/l
RT: 9.143 min Scan# 2779
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 93 Resp: 69071
Ion Ratio Lower Upper
93 100
95 84.5 0.0 169.0
174 102.9 0.0 205.8









#41
Bromodichloromethane
Concen: 18.87 ug/l
RT: 9.338 min Scan# 2852
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 83 Resp: 146494
Ion Ratio Lower Upper
83 100
85 63.2 50.6 75.8
127 8.6 6.9 10.3

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#42
Vinyl Acetate
Concen: 115.59 ug/l
RT: 5.793 min Scan# 1530
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 43 Resp: 1438940
Ion Ratio Lower Upper
43 100
86 8.4 6.7 10.1






#43
Ethyl Acetate
Concen: 23.87 ug/l
RT: 6.836 min Scan# 1919
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED

MMDadoda
3/30/2021 10:41:08 AM







Tgt Ion: 43 Resp: 150591
Ion Ratio Lower Upper
43 100
45 7.9 0.0 0.0#



#44
Isopropyl Acetate
Concen: 24.46 ug/l
RT: 8.089 min Scan# 2386
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 43 Resp: 266113
Ion Ratio Lower Upper
43 100
61 25.5 0.0 0.0#









#45
1,4-Dioxane
Concen: 387.91 ug/l
RT: 9.132 min Scan# 2775
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 88 Resp: 34087
Ion Ratio Lower Upper
88 100
43 33.7 27.0 40.4
58 72.4 57.9 86.9

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#46
Methyl methacrylate
Concen: 24.19 ug/l
RT: 9.135 min Scan# 2776
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 41 Resp: 122820
Ion Ratio Lower Upper
41 100
69 81.1 40.6 121.7
39 66.4 33.2 99.6









#47
n-amy1 Acetate
Concen: 15.63 ug/l
RT: 12.020 min Scan# 3852
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 43 Resp: 112430
Ion Ratio Lower Upper
43 100
55 22.1 17.7 26.5

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#48
t-1,3-Dichloropropene
Concen: 17.83 ug/l
RT: 10.323 min Scan# 3219
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 75 Resp: 158248
Ion Ratio Lower Upper
75 100
77 31.1 24.9 37.3









#49
cis-1,3-Dichloropropene
Concen: 18.47 ug/l
RT: 9.776 min Scan# 3015
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:	75	Resp:	170114
Ion	Ratio	Lower	Upper
75	100		
77	31.7	25.4	38.0
39	58.5	46.8	70.2

Instrument :
MSVOA_N
ClientSampled :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#50
1,1,2-Trichloroethane
Concen: 20.27 ug/l
RT: 10.502 min Scan# 3286
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Tgt Ion:	97	Resp:	100026
Ion	Ratio	Lower	Upper
97	100		
83	86.0	68.8	103.2
85	56.0	44.8	67.2
99	62.6	50.1	75.1





#51
Ethyl methacrylate
Concen: 20.69 ug/l
RT: 10.371 min Scan# 3237
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Tgt Ion: 69 Resp: 156358
Ion Ratio Lower Upper
69 100
41 69.7 34.8 104.5
39 45.0 22.5 67.5

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#52
1,3-Dichloropropane
Concen: 19.46 ug/l
RT: 10.647 min Scan# 3340
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52



Tgt Ion: 76 Resp: 166100
Ion Ratio Lower Upper
76 100
78 32.8 0.0 65.6









#53
Dibromochloromethane
Concen: 18.90 ug/l
RT: 10.843 min Scan# 3413
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:129 Resp: 112532
Ion Ratio Lower Upper
129 100
127 77.1 38.6 115.7

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#54
1,2-Dibromoethane
Concen: 19.91 ug/l
RT: 10.948 min Scan# 3452
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:107 Resp: 103188
Ion Ratio Lower Upper
107 100
109 94.7 75.8 113.6









#55
2-Chloroethyl vinyl ether
Concen: 52.77 ug/l
RT: 9.631 min Scan# 2961
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 63 Resp: 121867
Ion Ratio Lower Upper
63 100
65 32.5 26.0 39.0
106 25.8 20.6 31.0

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#56
Bromoform
Concen: 20.41 ug/l
RT: 12.071 min Scan# 3871
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 173 Resp: 82050
Ion Ratio Lower Upper
173 100
175 47.9 38.3 57.5
254 0.0 0.0 0.0









#57
Chlorobenzene-d5
Concen: 30.00 ug/l
RT: 11.350 min Scan# 3602
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 117 Resp: 400913
Ion Ratio Lower Upper
117 100
82 56.3 45.0 67.6
119 31.9 25.5 38.3

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#58
4-Methyl-2-Pentanone
Concen: 134.55 ug/l
RT: 9.920 min Scan# 3069
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 43 Resp: 791312
Ion Ratio Lower Upper
43 100
58 38.0 30.4 45.6
85 17.3 13.8 20.8









#59
2-Hexanone
Concen: 132.41 ug/l
RT: 10.693 min Scan# 3357
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 43 Resp: 544697
Ion Ratio Lower Upper
43 100
58 52.2 41.8 62.6
57 18.6 14.9 22.3

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#60
4-Bromofluorobenzene
Concen: 27.40 ug/l
RT: 12.345 min Scan# 3973
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 95 Resp: 187495
Ion Ratio Lower Upper
95 100
174 85.1 68.1 102.1
176 81.7 65.4 98.0









#61
Tetrachloroethene
Concen: 20.65 ug/l
RT: 10.572 min Scan# 3312
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:164 Resp: 108321
Ion Ratio Lower Upper
164 100
166 125.6 100.5 150.7
129 93.4 74.7 112.1
131 86.8 69.4 104.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#62
Toluene
Concen: 20.27 ug/l
RT: 10.092 min Scan# 3133
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 91 Resp: 429887
Ion Ratio Lower Upper
91 100
92 57.7 46.2 69.2









#63
Toluene-d8
Concen: 30.82 ug/l
RT: 10.028 min Scan# 3109
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 98 Resp: 557940
Ion Ratio Lower Upper
98 100
100 63.9 51.1 76.7

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#64
Chlorobenzene
Concen: 18.84 ug/l
RT: 11.377 min Scan# 3612
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 112 Resp: 255842
Ion Ratio Lower Upper
112 100
114 32.0 25.6 38.4









#65
1,1,1,2-Tetrachloroethane
Concen: 19.06 ug/l
RT: 11.452 min Scan# 3640
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:131 Resp: 99655
Ion Ratio Lower Upper
131 100
133 94.2 0.0 188.4
119 67.2 0.0 134.4

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#66
Ethyl Benzene
Concen: 19.35 ug/l
RT: 11.455 min Scan# 3641
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 91 Resp: 466172
Ion Ratio Lower Upper
91 100
106 31.9 25.5 38.3









#67
m/p-Xylenes
Concen: 37.29 ug/l
RT: 11.564 min Scan# 3682
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:106 Resp: 345684
Ion Ratio Lower Upper
106 100
91 206.0 164.8 247.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#68
o-Xylene
Concen: 18.77 ug/l
RT: 11.892 min Scan# 3804
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:106 Resp: 171062
Ion Ratio Lower Upper
106 100
91 214.3 107.1 321.4









#69
Styrene
Concen: 18.41 ug/l
RT: 11.908 min Scan# 3810
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:104 Resp: 275409
Ion Ratio Lower Upper
104 100
78 51.3 41.0 61.6
103 54.6 43.7 65.5

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#70
Isopropylbenzene
Concen: 18.95 ug/l
RT: 12.192 min Scan# 3916
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Tgt Ion:105 Resp: 452539
Ion Ratio Lower Upper
105 100
120 26.7 18.7 34.7





#71
1,1,2,2-Tetrachloroethane
Concen: 22.17 ug/l
RT: 12.447 min Scan# 4011
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Tgt Ion: 83 Resp: 148850
Ion Ratio Lower Upper
83 100
131 9.7 4.9 14.6
85 62.8 31.4 94.2

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#72
1,2,3-Trichloropropane
Concen: 21.60 ug/l m
RT: 12.498 min Scan# 4030
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Tgt Ion: 75 Resp: 144643
Ion Ratio Lower Upper
75 100
77 155.4 0.0 313.6





#73
Bromobenzene
Concen: 19.41 ug/l
RT: 12.468 min Scan# 4019
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Tgt Ion:156 Resp: 112514
Ion Ratio Lower Upper
156 100
77 199.8 0.0 399.6
158 96.4 0.0 192.8

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#74
n-propylbenzene
Concen: 18.83 ug/l
RT: 12.535 min Scan# 4044
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52



Tgt Ion: 91 Resp: 497879
Ion Ratio Lower Upper
91 100
120 22.6 0.0 45.2









#75
2-Chlorotoluene
Concen: 18.03 ug/l
RT: 12.618 min Scan# 4075
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 91 Resp: 301524
Ion Ratio Lower Upper
91 100
126 34.3 0.0 68.6

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#76
1,3,5-Trimethylbenzene
Concen: 17.53 ug/l
RT: 12.677 min Scan# 4097
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Tgt Ion: 105 Resp: 370009
Ion Ratio Lower Upper
105 100
120 49.1 0.0 98.2





#77
t-1,4-Dichloro-2-butene
Concen: 19.95 ug/l
RT: 12.246 min Scan# 3936
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Tgt Ion: 75 Resp: 46984
Ion Ratio Lower Upper
75 100
53 99.3 49.6 149.0
89 45.8 22.9 68.7

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#78
4-Chlorotoluene
Concen: 17.64 ug/l
RT: 12.718 min Scan# 4112
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52



Tgt Ion: 91 Resp: 299417
Ion Ratio Lower Upper
91 100
126 32.7 0.0 65.4









#79
tert-butylbenzene
Concen: 17.99 ug/l
RT: 12.938 min Scan# 4194
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:119 Resp: 316842
Ion Ratio Lower Upper
119 100
91 66.7 0.0 133.4
134 25.5 0.0 51.0

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#80
1,2,4-Trimethylbenzene
Concen: 17.13 ug/l
RT: 12.983 min Scan# 4211
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52


Tgt Ion:105 Resp: 354677
Ion Ratio Lower Upper
105 100
120 45.3 0.0 90.6





#81
sec-Butylbenzene
Concen: 18.46 ug/l
RT: 13.115 min Scan# 4260
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Tgt Ion:105 Resp: 410254
Ion Ratio Lower Upper
105 100
134 20.0 0.0 40.0

Manual Integrations
APPROVED



MMDadoda
3/30/2021 10:41:08 AM




#82
p-Isopropyltoluene
Concen: 17.21 ug/l
RT: 13.233 min Scan# 4304
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Tgt Ion:119 Resp: 358212
Ion Ratio Lower Upper
119 100
134 26.7 0.0 53.4
91 23.8 0.0 47.6









#83
1,3-Dichlorobenzene
Concen: 18.24 ug/l
RT: 13.227 min Scan# 4302
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52





Tgt Ion:146 Resp: 181171
Ion Ratio Lower Upper
146 100
111 38.6 19.3 57.9
148 63.9 31.9 95.9



#84
1,4-Dichlorobenzene
Concen: 18.52 ug/l
RT: 13.308 min Scan# 4332
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52




Tgt Ion:146 Resp: 178841
Ion Ratio Lower Upper
146 100
111 36.7 18.4 55.0
148 63.7 31.9 95.5



Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED

MMDadoda
3/30/2021 10:41:08 AM







#85
n-Butylbenzene
Concen: 17.60 ug/l
RT: 13.560 min Scan# 4426
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 91 Resp: 293028
Ion Ratio Lower Upper
91 100
92 51.3 0.0 102.6
134 24.5 0.0 49.0

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#86
Hexachloroethane
Concen: 19.71 ug/l
RT: 13.815 min Scan# 4521
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 117 Resp: 66901
Ion Ratio Lower Upper
117 100
201 86.5 43.3 129.8









#87
1,2-Dichlorobenzene
Concen: 18.38 ug/l
RT: 13.597 min Scan# 4440
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:146 Resp: 180093
Ion Ratio Lower Upper
146 100
111 40.6 20.3 60.9
148 64.0 32.0 96.0

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#88
1,2-Dibromo-3-Chloropropane
Concen: 19.36 ug/l
RT: 14.214 min Scan# 4670
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion: 75 Resp: 30412
Ion Ratio Lower Upper
75 100
155 85.4 68.3 102.5
157 109.7 87.8 131.6









#89
1,2,4-Trichlorobenzene
Concen: 20.08 ug/l
RT: 14.850 min Scan# 4907
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:180 Resp: 111598
Ion Ratio Lower Upper
180 100
182 92.9 74.3 111.5
145 29.2 23.4 35.0

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020




Manual Integrations
APPROVED


MMDadoda
3/30/2021 10:41:08 AM



#90
Hexachlorobutadiene
Concen: 23.46 ug/l
RT: 14.949 min Scan# 4944
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:225 Resp: 59897
Ion Ratio Lower Upper
225 100
223 62.3 0.0 124.6
227 63.3 0.0 126.6









#91
Naphthalene
Concen: 21.16 ug/l
RT: 15.070 min Scan# 4989
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:128 Resp: 346332
Ion Ratio Lower Upper
128 100
127 12.7 10.2 15.2
129 10.8 8.6 13.0

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Manual Integrations
APPROVED

MMDadoda
3/30/2021 10:41:08 AM



#92
1,2,3-Trichlorobenzene
Concen: 19.81 ug/l
RT: 15.249 min Scan# 5056
Delta R.T. 0.000 min
Lab File: VN066311.D
Acq: 24 Mar 2021 14:52

Tgt Ion:180 Resp: 110621
Ion Ratio Lower Upper
180 100
182 95.0 0.0 190.0
145 30.4 0.0 60.8

