



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032421\
 Data File : VN066308.D
 Acq On : 24 Mar 2021 13:20
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu Mar 25 05:52:50 2021



AutoFind: Scans 3973, 3974, 3975; Background Corrected with Scan 3957

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	13631	PASS
75	95	30	60	51.4	33000	PASS
95	95	100	100	100.0	64179	PASS
96	95	5	9	6.6	4245	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	83.4	53555	PASS
175	174	5	9	7.4	3987	PASS
176	174	95	101	98.2	52573	PASS
177	176	5	9	6.0	3155	PASS