

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
 Data File : VN086066.D
 Acq On : 24 Mar 2025 12:38
 Operator : JC\MD
 Sample : VN0324MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0324MBL01

Quant Time: Mar 25 01:20:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	141819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	249369	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	226919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	90093	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	110982	57.149	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.300%
35) Dibromofluoromethane	8.165	113	95862	55.075	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.160%
50) Toluene-d8	10.565	98	307439	48.668	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.340%
62) 4-Bromofluorobenzene	12.847	95	90718	40.268	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	80.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
Data File : VN086066.D
Acq On : 24 Mar 2025 12:38
Operator : JC\MD
Sample : VN0324MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0324MBL01

Quant Time: Mar 25 01:20:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

