

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
 Data File : VN086072.D
 Acq On : 24 Mar 2025 15:12
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 25 01:24:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	190367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	346741	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316044	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	142900	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	135242	51.881	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.760%
35) Dibromofluoromethane	8.165	113	118367	48.908	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.820%
50) Toluene-d8	10.565	98	415345	47.286	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.580%
62) 4-Bromofluorobenzene	12.847	95	148289	47.339	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
Data File : VN086072.D
Acq On : 24 Mar 2025 15:12
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 25 01:24:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

