

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
 Data File : VN086080.D
 Acq On : 24 Mar 2025 18:25
 Operator : JC\MD
 Sample : Q1619-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Mar 25 01:29:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	281256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	515188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	492000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	212760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	194148	50.410	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.820%	
35) Dibromofluoromethane	8.165	113	171725	47.755	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.520%	
50) Toluene-d8	10.565	98	648577	49.696	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.400%	
62) 4-Bromofluorobenzene	12.847	95	218513	46.949	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 93.900%	
Target Compounds						Qvalue
16) Acetone	4.430	43	50378	42.816	ug/l	94
20) Methylene Chloride	5.277	84	17549	5.098	ug/l	88
43) Isopropyl Acetate	8.682	43	153518	17.735	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
Data File : VN086080.D
Acq On : 24 Mar 2025 18:25
Operator : JC\MD
Sample : Q1619-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Mar 25 01:29:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

