

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
 Data File : VN086085.D
 Acq On : 24 Mar 2025 20:26
 Operator : JC\MD
 Sample : Q1622-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Field-Blank

Quant Time: Mar 25 01:32:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	262167	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	483807	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	422031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	173343	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	184906	51.506	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.020%
35) Dibromofluoromethane	8.165	113	163860	48.524	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.040%
50) Toluene-d8	10.565	98	562711	45.914	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.820%
62) 4-Bromofluorobenzene	12.847	95	187763	42.958	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	85.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
Data File : VN086085.D
Acq On : 24 Mar 2025 20:26
Operator : JC\MD
Sample : Q1622-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Field-Blank

Quant Time: Mar 25 01:32:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

