

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
 Data File : VN086086.D
 Acq On : 24 Mar 2025 20:50
 Operator : JC\MD
 Sample : Q1622-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW1

Quant Time: Mar 25 01:32:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	252685	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	462481	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	398785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	163397	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	178605	51.618	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	103.240%	
35) Dibromofluoromethane	8.165	113	158567	49.121	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.240%	
50) Toluene-d8	10.565	98	540982	46.176	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	92.360%	
62) 4-Bromofluorobenzene	12.847	95	178180	42.646	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	85.300%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	2766	2.617	ug/l	93
30) Chloroform	7.965	83	4502	0.805	ug/l #	81
44) Trichloroethene	9.347	130	1076	0.311	ug/l	76
64) Tetrachloroethene	11.094	164	1731	0.544	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
Data File : VN086086.D
Acq On : 24 Mar 2025 20:50
Operator : JC\MD
Sample : Q1622-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW1

Quant Time: Mar 25 01:32:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

