

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
 Data File : VN086087.D
 Acq On : 24 Mar 2025 21:14
 Operator : JC\MD
 Sample : Q1622-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW3

Quant Time: Mar 25 01:33:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	243235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	452406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	398702	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	157433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	173567	52.111	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.220%
35) Dibromofluoromethane	8.165	113	153852	48.722	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	10.565	98	532835	46.494	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.980%
62) 4-Bromofluorobenzene	12.847	95	177536	43.438	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.880%
Target Compounds						
16) Acetone	4.436	43	3370	3.312	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032425\
Data File : VN086087.D
Acq On : 24 Mar 2025 21:14
Operator : JC\MD
Sample : Q1622-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW3

Quant Time: Mar 25 01:33:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

