

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032519\
 Data File : VN054661.D
 Acq On : 25 Mar 2019 10:11
 Operator : JC/SP
 Sample : VN0325WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0325WBL01

Quant Time: Mar 26 01:49:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	588038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	904767	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	763258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	274896	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	362339	49.54	ug/l	0.00
Spiked Amount			Recovery	=	99.08%	
35) Dibromofluoromethane	7.59	113	280850	48.18	ug/l	0.00
Spiked Amount			Recovery	=	96.36%	
50) Toluene-d8	10.09	98	1079654	49.27	ug/l	0.00
Spiked Amount			Recovery	=	98.54%	
62) 4-Bromofluorobenzene	12.40	95	337619	45.28	ug/l	0.00
Spiked Amount			Recovery	=	90.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032519\
Data File : VN054661.D
Acq On : 25 Mar 2019 10:11
Operator : JC/SP
Sample : VN0325WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0325WBL01

Quant Time: Mar 26 01:49:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

