

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032519\
 Data File : VN054677.D
 Acq On : 25 Mar 2019 17:11
 Operator : JC/SP
 Sample : K2071-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-1

Quant Time: Mar 26 03:38:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	564824	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	901538	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	748400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	280814	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	371140	52.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.66%	
35) Dibromofluoromethane	7.59	113	292630	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
50) Toluene-d8	10.09	98	1078708	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
62) 4-Bromofluorobenzene	12.40	95	331808	44.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.32%	
Target Compounds						
20) Methylene Chloride	4.55	84	17459	2.804	ug/l #	83
43) Isopropyl Acetate	8.18	43	10929	1.079	ug/l #	93
84) 1,2,4-Trimethylbenzene	13.04	105	30207	1.660	ug/l	99
95) Naphthalene	15.13	128	25775	2.573	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032519\
Data File : VN054677.D
Acq On : 25 Mar 2019 17:11
Operator : JC/SP
Sample : K2071-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-1

Quant Time: Mar 26 03:38:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

