

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032520\
 Data File : VN060698.D
 Acq On : 25 Mar 2020 15:36
 Operator : JC/MD
 Sample : L2008-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-4

Quant Time: Mar 26 05:08:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	200896	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	332909	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	313295	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147338	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130510	47.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.64%	
35) Dibromofluoromethane	7.56	113	99419	49.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.78%	
50) Toluene-d8	10.08	98	410524	49.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.86%	
62) 4-Bromofluorobenzene	12.40	95	155316	51.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.08%	
Target Compounds						
16) Acetone	3.78	43	3560	3.644	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032520\
Data File : VN060698.D
Acq On : 25 Mar 2020 15:36
Operator : JC/MD
Sample : L2008-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-4

Quant Time: Mar 26 05:08:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

