

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032521\
 Data File : VN066327.D
 Acq On : 25 Mar 2021 12:44
 Operator : JC/MD
 Sample : M1741-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OF-1

Manual Integrations
 APPROVED

SAM
 3/29/2021 5:11:54 PM

Quant Time: Mar 26 04:39:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 25 05:52:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.102	128	131671	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.512	114	702743	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.347	117	598695	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.947	65	298554	29.50	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	98.33%	
60) 4-Bromofluorobenzene	12.345	95	261443	27.98	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	93.27%	
63) Toluene-d8	10.028	98	834501	29.70	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	99.00%	
Target Compounds						Qvalue
15) Acetone	3.731	58	22235m	17.54	ug/l	
30) 2-Butanone	6.753	43	10771m	1.75	ug/l	
58) 4-Methyl-2-Pentanone	9.923	43	32993	2.75	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032521\
Data File : VN066327.D
Acq On : 25 Mar 2021 12:44
Operator : JC/MD
Sample : M1741-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OF-1

Quant Time: Mar 26 04:39:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 25 05:52:50 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

SAM
3/29/2021 5:11:54 PM

