


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032521\
 Data File : VN066321.D
 Acq On : 25 Mar 2021 8:40
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Thu Mar 25 05:52:50 2021



AutoFind: Averaged scan 3959 to 3998; Bkg corrected with scan 3958

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	2521	PASS
75	95	30	60	50.8	6112	PASS
95	95	100	100	100.0	12038	PASS
96	95	5	9	7.5	908	PASS
173	174	0.00	2	0.6	64	PASS
174	95	50	100	84.4	10162	PASS
175	174	5	9	7.5	763	PASS
176	174	95	101	96.5	9802	PASS
177	176	5	9	8.0	781	PASS