

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071538.D
 Acq On : 25 Mar 2022 11:06
 Operator : JC\MD
 Sample : VN0325WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0325WBL01

Quant Time: Mar 26 03:57:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	667802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1099689	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1058563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	362688	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	391172	47.148	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.300%
35) Dibromofluoromethane	8.016	113	318451	46.077	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.160%
50) Toluene-d8	10.439	98	1343769	48.204	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.400%
62) 4-Bromofluorobenzene	12.727	95	427350	46.786	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071538.D
Acq On : 25 Mar 2022 11:06
Operator : JC\MD
Sample : VN0325WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0325WBL01

Quant Time: Mar 26 03:57:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

