

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071541.D
 Acq On : 25 Mar 2022 12:40
 Operator : JC\MD
 Sample : PB143557TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143557TB

Quant Time: Mar 26 03:59:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	747024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1237426	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1226686	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	441873	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	432009	46.548	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.100%
35) Dibromofluoromethane	8.022	113	357837	46.013	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.020%
50) Toluene-d8	10.439	98	1556711	49.627	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.260%
62) 4-Bromofluorobenzene	12.727	95	507111	49.339	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.680%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	26796	7.044	ug/l	96
20) Methylene Chloride	5.098	84	12251	1.382	ug/l #	85
43) Isopropyl Acetate	8.563	43	68634	3.131	ug/l #	91
95) Naphthalene	15.498	128	16399	3.897	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071541.D
Acq On : 25 Mar 2022 12:40
Operator : JC\MD
Sample : PB143557TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143557TB

Quant Time: Mar 26 03:59:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

