

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071543.D
 Acq On : 25 Mar 2022 13:28
 Operator : JC\MD
 Sample : PB143557ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143557ZHE#02

Quant Time: Mar 26 04:00:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	764792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1308564	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1293150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	485327	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	448579	47.210	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.420%
35) Dibromofluoromethane	8.016	113	369774	44.963	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.920%
50) Toluene-d8	10.439	98	1631502	49.184	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.360%
62) 4-Bromofluorobenzene	12.727	95	555746	51.131	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.260%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.828	74	14979	2.893	ug/l	95
16) Acetone	4.292	43	26664	6.846	ug/l	98
20) Methylene Chloride	5.092	84	14482	1.596	ug/l	97
43) Isopropyl Acetate	8.557	43	89296	3.947	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071543.D
Acq On : 25 Mar 2022 13:28
Operator : JC\MD
Sample : PB143557ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143557ZHE#02

Quant Time: Mar 26 04:00:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

