

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071546.D
 Acq On : 25 Mar 2022 14:41
 Operator : JC\MD
 Sample : PB143557ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143557ZHE#05

Quant Time: Mar 26 04:01:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	742114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1286037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1270993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	471779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	447293	48.513	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.020%
35) Dibromofluoromethane	8.022	113	366161	45.304	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.600%
50) Toluene-d8	10.439	98	1593606	48.883	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.760%
62) 4-Bromofluorobenzene	12.727	95	535088	50.093	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.180%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.828	74	15527	3.090	ug/l	98
16) Acetone	4.293	43	26227	6.940	ug/l #	88
20) Methylene Chloride	5.098	84	13335	1.514	ug/l	94
43) Isopropyl Acetate	8.563	43	134009	6.243	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071546.D
Acq On : 25 Mar 2022 14:41
Operator : JC\MD
Sample : PB143557ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143557ZHE#05

Quant Time: Mar 26 04:01:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

