

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071547.D
 Acq On : 25 Mar 2022 15:05
 Operator : JC\MD
 Sample : PB143557ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143557ZHE#06

Quant Time: Mar 26 04:02:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	776748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1325178	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1301550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	471591	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	454672	47.115	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.220%
35) Dibromofluoromethane	8.022	113	376733	45.235	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.480%
50) Toluene-d8	10.439	98	1668767	49.677	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.360%
62) 4-Bromofluorobenzene	12.727	95	549079	49.884	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.760%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	27186	6.873	ug/l	93
20) Methylene Chloride	5.098	84	14439	1.567	ug/l	97
43) Isopropyl Acetate	8.563	43	135923	6.139	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071547.D
Acq On : 25 Mar 2022 15:05
Operator : JC\MD
Sample : PB143557ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143557ZHE#06

Quant Time: Mar 26 04:02:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

