

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071548.D
 Acq On : 25 Mar 2022 15:29
 Operator : JC\MD
 Sample : PB143557ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143557ZHE#07

Quant Time: Mar 26 04:02:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	739874	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1251171	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1236269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	441821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	437163	47.558	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.120%
35) Dibromofluoromethane	8.016	113	357462	45.460	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.920%
50) Toluene-d8	10.439	98	1570142	49.505	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.020%
62) 4-Bromofluorobenzene	12.727	95	514262	49.485	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.960%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	27357	7.261	ug/l	96
20) Methylene Chloride	5.093	84	12759	1.453	ug/l	85
43) Isopropyl Acetate	8.563	43	148911	7.189	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071548.D
Acq On : 25 Mar 2022 15:29
Operator : JC\MD
Sample : PB143557ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143557ZHE#07

Quant Time: Mar 26 04:02:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

