

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071551.D
 Acq On : 25 Mar 2022 16:42
 Operator : JC\MD
 Sample : PB143557ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143557ZHE#10

Quant Time: Mar 26 04:03:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	815210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1398279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1377764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	504361	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	476224	47.020	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.040%
35) Dibromofluoromethane	8.022	113	394094	44.846	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.700%
50) Toluene-d8	10.439	98	1749027	49.344	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.680%
62) 4-Bromofluorobenzene	12.727	95	578855	49.840	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.680%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	28314	6.820	ug/l	95
18) Methyl Acetate	4.857	43	24959	1.077	ug/l	97
20) Methylene Chloride	5.098	84	11423	1.181	ug/l #	95
43) Isopropyl Acetate	8.569	43	232332	10.199	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071551.D
Acq On : 25 Mar 2022 16:42
Operator : JC\MD
Sample : PB143557ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143557ZHE#10

Quant Time: Mar 26 04:03:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

