

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071552.D
 Acq On : 25 Mar 2022 17:06
 Operator : JC\MD
 Sample : PB143557ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143557ZHE#11

Quant Time: Mar 26 04:04:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	812861	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1388692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1371398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	515409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	477354	47.268	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.540%		
35) Dibromofluoromethane	8.022	113	393127	45.045	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	90.080%		
50) Toluene-d8	10.439	98	1740155	49.433	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.860%		
62) 4-Bromofluorobenzene	12.727	95	576157	49.950	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	99.900%		
Target Compounds						Qvalue
8) Diethyl Ether	3.828	74	14012	2.546	ug/l	86
16) Acetone	4.298	43	27962	6.755	ug/l	98
18) Methyl Acetate	4.857	43	25630	1.147	ug/l	97
20) Methylene Chloride	5.098	84	12966	1.344	ug/l	89
43) Isopropyl Acetate	8.569	43	218105	9.618	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071552.D
Acq On : 25 Mar 2022 17:06
Operator : JC\MD
Sample : PB143557ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143557ZHE#11

Quant Time: Mar 26 04:04:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

