

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071553.D
 Acq On : 25 Mar 2022 17:30
 Operator : JC\MD
 Sample : N2060-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-20220321

Quant Time: Mar 26 04:04:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	797167	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1363234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1325107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	482191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	465190	46.970	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.940%
35) Dibromofluoromethane	8.016	113	384527	44.882	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.760%
50) Toluene-d8	10.439	98	1697032	49.108	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.220%
62) 4-Bromofluorobenzene	12.727	95	552930	48.832	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.660%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.834	74	14311	2.652	ug/l	94
16) Acetone	4.293	43	28751	7.082	ug/l #	83
20) Methylene Chloride	5.087	84	14080	1.489	ug/l	91
43) Isopropyl Acetate	8.563	43	165119	7.324	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071553.D
Acq On : 25 Mar 2022 17:30
Operator : JC\MD
Sample : N2060-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-20220321

Quant Time: Mar 26 04:04:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

