

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071554.D
 Acq On : 25 Mar 2022 17:55
 Operator : JC\MD
 Sample : N2077-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-4(2.5-3.0)

Quant Time: Mar 26 04:05:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	785026	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1328181	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1330676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	494054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	459461	47.109	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.220%
35) Dibromofluoromethane	8.016	113	378651	45.363	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.720%
50) Toluene-d8	10.439	98	1681520	49.943	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.880%
62) 4-Bromofluorobenzene	12.727	95	560271	50.786	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.580%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.828	74	13518	2.543	ug/l	88
16) Acetone	4.293	43	41824	10.462	ug/l #	88
20) Methylene Chloride	5.087	84	18525	1.989	ug/l	89
43) Isopropyl Acetate	8.563	43	127606	5.724	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071554.D
Acq On : 25 Mar 2022 17:55
Operator : JC\MD
Sample : N2077-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-4(2.5-3.0)

Quant Time: Mar 26 04:05:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

