

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071555.D
 Acq On : 25 Mar 2022 18:19
 Operator : JC\MD
 Sample : N2077-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-4

Quant Time: Mar 26 04:05:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	763148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1298442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1253164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	473398	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	450149	47.477	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.960%		
35) Dibromofluoromethane	8.022	113	371653	45.544	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.080%		
50) Toluene-d8	10.439	98	1607027	48.824	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.640%		
62) 4-Bromofluorobenzene	12.727	95	532987	49.420	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.840%		
Target Compounds						Qvalue
8) Diethyl Ether	3.828	74	13105	2.536	ug/l	94
16) Acetone	4.287	43	55723	14.338	ug/l	98
20) Methylene Chloride	5.104	84	17196	1.899	ug/l	95
39) Methylcyclohexane	9.457	83	21042	1.458	ug/l	93
43) Isopropyl Acetate	8.563	43	198347	9.344	ug/l #	67
69) o-Xylene	12.274	106	32983	1.901	ug/l	100
84) 1,2,4-Trimethylbenzene	13.362	105	148968	4.686	ug/l	99
95) Naphthalene	15.504	128	14244	3.768	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071555.D
Acq On : 25 Mar 2022 18:19
Operator : JC\MD
Sample : N2077-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-4

Quant Time: Mar 26 04:05:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

