

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071558.D
 Acq On : 25 Mar 2022 19:32
 Operator : JC\MD
 Sample : N2077-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-3-2(3.0-3.5)

Quant Time: Mar 26 04:07:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	756605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1295914	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1271959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	450342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	442898	47.117	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.240%		
35) Dibromofluoromethane	8.022	113	364342	44.735	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	89.480%		
50) Toluene-d8	10.439	98	1609896	49.006	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.020%		
62) 4-Bromofluorobenzene	12.727	95	521321	48.432	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.860%		
Target Compounds						
8) Diethyl Ether	3.834	74	13616	2.658	ug/l	91
16) Acetone	4.287	43	71213	18.482	ug/l	98
43) Isopropyl Acetate	8.557	43	87831	3.917	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071558.D
Acq On : 25 Mar 2022 19:32
Operator : JC\MD
Sample : N2077-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-3-2-(3.0-3.5)

Quant Time: Mar 26 04:07:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

