

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071559.D
 Acq On : 25 Mar 2022 19:56
 Operator : JC\MD
 Sample : N2077-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-3-2

Quant Time: Mar 26 04:07:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 25 09:52:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	788457	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1333086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1341441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	566138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	461075	47.069	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.140%		
35) Dibromofluoromethane	8.016	113	378745	45.207	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	90.420%		
50) Toluene-d8	10.439	98	1673622	49.526	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.060%		
62) 4-Bromofluorobenzene	12.727	95	602624	54.424	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	108.840%		
Target Compounds						Qvalue
8) Diethyl Ether	3.822	74	13140	2.462	ug/l	97
16) Acetone	4.287	43	69711	17.362	ug/l	98
20) Methylene Chloride	5.092	84	117139	12.520	ug/l	97
43) Isopropyl Acetate	8.557	43	107588	4.743	ug/l #	77
73) Isopropylbenzene	12.574	105	1245990	24.963	ug/l	100
78) n-propylbenzene	12.915	91	1044082	19.661	ug/l	100
83) tert-Butylbenzene	13.321	119	98785	2.777	ug/l	93
85) sec-Butylbenzene	13.498	105	420140	8.912	ug/l #	77
89) n-Butylbenzene	13.939	91	81314	2.825	ug/l	87
95) Naphthalene	15.498	128	18729	3.823	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
Data File : VN071559.D
Acq On : 25 Mar 2022 19:56
Operator : JC\MD
Sample : N2077-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A-3-2

Quant Time: Mar 26 04:07:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 25 09:52:40 2022
Response via : Initial Calibration

