


Chloromethane

Response Ratio



$$\text{Response} = 5.655\text{e-}001 * \text{Amt} + 2.073\text{e-}002$$


Coef of Det (r^2) = 0.997467 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M

Calibration Table Last Updated: Fri Mar 25 09:52:40 2022

Methyl Acetate

Response Ratio



$$\text{Response} = 6.507\text{e-}001 * \text{Amt} + 1.660\text{e-}002$$


Coef of Det (r^2) = 0.997569 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M

Calibration Table Last Updated: Fri Mar 25 09:52:40 2022

Isopropyl Acetate

Response Ratio



$$\text{Response} = 7.830\text{e-}001 * \text{Amt} + 6.433\text{e-}003$$


Coef of Det (r^2) = 0.997525 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M

Calibration Table Last Updated: Fri Mar 25 09:52:40 2022

2-Chloroethyl Vinyl ether

Response Ratio



$$\text{Response} = 1.840\text{e-}001 * \text{Amt} - 6.877\text{e-}002$$


Coef of Det (r^2) = 0.997994 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M

Calibration Table Last Updated: Fri Mar 25 09:52:40 2022

1,1,2,2-Tetrachloroethane


Response Ratio



Response = 1.213e+000 * Amt + 9.325e-002
 Coef of Det (r^2) = 0.992825 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Calibration Table Last Updated: Fri Mar 25 09:52:40 2022

Naphthalene

Response Ratio



$$\text{Response} = 2.720\text{e}+000 * \text{Amt} - 1.748\text{e}-001$$

Coef of Det (r^2) = 0.997395 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M

Calibration Table Last Updated: Fri Mar 25 09:52:40 2022