



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032522\
 Data File : VN071526.D
 Acq On : 24 Mar 2022 17:35
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032522W.M
 Title : SW846 8260
 Last Update : Fri Mar 25 09:52:40 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	15840	PASS
75	95	30	60	48.3	42963	PASS
95	95	100	100	100.0	88909	PASS
96	95	5	9	7.2	6432	PASS
173	174	0.00	2	1.0	777	PASS
174	95	50	100	85.9	76368	PASS
175	174	5	9	7.2	5466	PASS
176	174	95	101	95.8	73144	PASS
177	176	5	9	6.3	4623	PASS