

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032523\
 Data File : VN077194.D
 Acq On : 25 Mar 2023 00:07
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Mar 27 02:49:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 27 02:47:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	456667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	779439	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	669251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	244149	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	334354	46.827	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.660%
35) Dibromofluoromethane	8.172	113	251867	47.622	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.240%
50) Toluene-d8	10.571	98	940247	47.445	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.900%
62) 4-Bromofluorobenzene	12.854	95	280283	43.094	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032523\
Data File : VN077194.D
Acq On : 25 Mar 2023 00:07
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Mar 27 02:49:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032523W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 27 02:47:06 2023
Response via : Initial Calibration

