

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032524\
 Data File : VN081560.D
 Acq On : 25 Mar 2024 14:30
 Operator : JC\MD
 Sample : P1838-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE137-SP300D-20240321

Quant Time: Mar 26 01:19:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	441439	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	769822	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	673621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	262357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	275084	48.865	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.720%
35) Dibromofluoromethane	8.165	113	241680	49.745	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.500%
50) Toluene-d8	10.565	98	905493	51.559	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.120%
62) 4-Bromofluorobenzene	12.847	95	278070	44.935	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	89.860%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.377	101	30172	5.365	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032524\
Data File : VN081560.D
Acq On : 25 Mar 2024 14:30
Operator : JC\MD
Sample : P1838-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE137-SP300D-20240321

Quant Time: Mar 26 01:19:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

