

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032524\
 Data File : VN081569.D
 Acq On : 25 Mar 2024 18:28
 Operator : JC\MD
 Sample : P1840-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 26 01:22:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	397737	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	711187	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	636154	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	248011	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	262915	51.834	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.660%
35) Dibromofluoromethane	8.165	113	226822	50.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.080%
50) Toluene-d8	10.565	98	838700	51.693	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.380%
62) 4-Bromofluorobenzene	12.853	95	263803	46.144	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	92.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032524\
Data File : VN081569.D
Acq On : 25 Mar 2024 18:28
Operator : JC\MD
Sample : P1840-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 26 01:22:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

